

Applicant



Certificate of Conformity



Number CoC-1902605-01 Project number 1902605 Page 1 of 3

Issued by : NMi Certin B.V.

Hugo de Grootplein 1 314 EG Dordrecht The Netherlands

: Janitza electronics GmbH

Vor dem Polstück 6 35633 Lahnau Germany

Submitted : A meter embedding IEC 61000-4-30 class A Power Quality functions

Manufacturer : Janitza

Type : UMG 512-PRO

Characteristics : See page 2 and further

In accordance with : IEC 61000-4-30 Ed. 3 (2015)

"Electromagnetic Compatibility (EMC) – Part 4-30: Testing and measurement techniques – Power quality measurement methods"

IEC 62586-2 Ed. 2 (2017)

"Power quality measurement in power supply systems - Part 2: Functional

tests and uncertainty requirements"

Measurement class : IEC 61000-4-30 class A

The undersigned declares that the described product is tested according to the above mentioned standard and meet their requirements, based on a non-recurrent examination. The appertaining test data is presented in type evaluation report number NMi-1902605-01, granted by NMi Certin B.V.

NMi Certin B.V. 25 April 2019

C. Oosterman

Head Certification Board

NMi Certin B.V. Thijsseweg 11 2629 JA Delft

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document only is permitted.









Certificate of Conformity

Number CoC-1902605-01 Project number 1902605 Page 2 of 3

IEC 61000-4-30 Power Quality functions tested

The following IEC 61000-4-30 measurement methods have been tested

Table 1 IEC 61000-4-30 Power Quality functions tested

IEC 62586-2 Clause	Parameter	IEC 61000-4-30 class	Comments
6.1 / 7.1	Power frequency	Α	50 Hz and 60 Hz
6.2 / 7.2	Magnitude of supply voltage	Α	230 V and 57,7 V
6.3 / 7.3	Flicker	Α	Class F3: 230 V, 60 Hz 120 V, 50 Hz
6.4 / 7.4	Supply voltage interruptions, dips and swells	A	50 and 60 Hz
6.5 / 7.5	Supply voltage unbalance	Α	
6.6 / 7.6	Voltage harmonics	A	
6.7 / 7.7	Voltage interharmonics	Α	
6.8 / 7.8	Mains signalling voltages on the voltage supply	Α	Method 2
6.9 / 7.9	Measurement of underdeviation and overdeviation parameters	Α	
6.10 / 7.10	Flagging	Α	
6.11 / 7.11	Clock uncertainty testing	Α	
6.12 / 7.12	Variation of external influence quantities	Α	Temperature: -10°C +55°C
6.13 / 7.13	Rapid Voltage Changes (RVC)	Α	
6.14 / 7.14	Magnitude of current	Α	
6.15 / 7.15	Harmonic current	A	
6.16 / 7.16	Interharmonic currents	Α	
6.17 / 7.17	Current unbalance	Α	
8	Calculation of measurement uncertainty and operating uncertainty	A	

A : compliance with class AS : compliance with class S--- : Not implemented

The tests are performed in accordance with IEC 62586-2 edition 2 (2017).



Certificate of Conformity

Number CoC-1902605-01 Project number 1902605 Page 3 of 3

Characteristics of the measuring instrument

In Table 2 the general characteristics of the measuring instrument are presented.

Table 2 General characteristics

Model	UMG 512-PRO	
U_{din}	230 V _{LN} or 57,7 V _{LN}	
I _{nom}	3 A	
f_{nom}	50 Hz and 60 Hz	
Temperature	Rated range of operation: -10°C to +55°C	
Power supply range	95 240 VAC 50/60 Hz 48 110 VDC (52.17.011) 48 110 VAC 50/60 Hz 24 150 VDC (52.17.003)	
Software version	V 5.011	
Hardware version	100	
Environmental application	Fixed (F), Indoor (I)	